

Radon Dosimetry Team RCE Chilton Didcot OX11 0RQ T +44 (0)1235 825373 www.gov.uk/ukhsa

## **CERTIFICATE OF PARTICIPATION**

**Radongas srl** participated in the 2022 UKHSA Intercomparison of Passive Radon Detectors with two sets of detectors.

The submitted results for set **136-1**, % measurement errors and corresponding ranks are given in Table 1 below for the five exposures:

Table 1 – set 136-1 - etched track detectors

Reference exposure [kBq m <sup>-3</sup> h]	Mean [kBq m <sup>-3</sup> h]	1SD [kBq m <sup>-3</sup> h]	% Measurement error	Rank
138	147.9	2.5	7.4	Α
383	387.3	10.4	2.9	Α
732	746.7	20.6	3.4	Α
1387	1382.4	41.6	3.0	А
2501	2478.5	61.0	2.6	Α

The submitted results for set **136-2**, % measurement errors and corresponding ranks are given in Table 2 below for the five exposures:

Table 2 – set 136-2 - badge (etched track detectors without rigid casing)

Reference exposure [kBq m <sup>-3</sup> h]	Mean [kBq m <sup>-3</sup> h]	1SD [kBq m <sup>-3</sup> h]	% Measurement error	Rank
138	142.0	3.7	3.9	Α
383	383.9	8.3	2.2	А
732	720.0	16.7	2.8	Α
1387	1395.7	21.7	1.7	А
2501	2499.7	82.5	3.3	А

% Measurement error=
$$\sqrt[2]{\left[\frac{Mean-Reference}{Reference}\times 100\right]^2+\left[\frac{1SD}{Mean}\times 100\right]^2}$$

## If the measurement error is:

- less than 10%, the rank is A
- greater than or equal to 10% and less than 20%, the rank is B
- greater than or equal to 20% and less than 30%, the rank is C
- greater than or equal to 30% and less than 40%, the rank is D
- greater than or equal to 40% and less than 50%, the rank is E
- greater than or equal to 50%, the rank is F

Cathere Milk

Catherine Miller 04/09/2023